



**Australian Government**  
**Department of Health**  
Therapeutic Goods Administration

## Public Summary

**Summary for ARTG Entry:** 356312 ACRA Regulatory Services Pty Ltd - Nerve-locating probe, single-use

**ARTG entry for** Medical Device Included Class IIa  
**Sponsor** ACRA Regulatory Services Pty Ltd  
**Postal Address** 7/84 Poinciana Avenue, Tewantin, QLD, 4565  
Australia  
**ARTG Start Date** 5/03/2021  
**Product Category** Medical Device Class IIa  
**Status** Active  
**Approval Area** Medical Devices

### Conditions

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.  
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

### Manufacturers

Name	Address
Technomed Europe	Amerikalaan 71 Maastricht Airport, , 6199 AE Netherlands

### Products

#### 1 . Nerve-locating probe, single-use

<b>Product Type</b>	Single Device Product	<b>Effective Date</b>	5/03/2021
<b>GMDN</b>	61089 Nerve-locating probe, single-use		
<b>Intended Purpose</b>	Single use sterile probes and surgical instruments consisting of insulated electrical conductor which have exposed tip or the outer cannula for stimulating/recording of bio potential signals, are intended to be used during surgery to detect or stimulate nerve action potentials and for tissue dissection.		

### Specific Conditions

No Specific Conditions included on Record

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/website-copyright.htm>.

Public Summary