



Australian Government
Department of Health
Therapeutic Goods Administration

Public Summary

Summary for ARTG Entry: 345835 Symbiotic Devices Pty Ltd - Adhesive, <specify>

ARTG entry for Medical Device Included Class 1
Sponsor Symbiotic Devices Pty Ltd
Postal Address Suite 138 / 245 Thomas Street, DANDENONG, VIC, 3175
Australia
ARTG Start Date 13/10/2020
Product Category Medical Device Class 1
Status Active
Approval Area Medical Devices

Conditions

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Manufacturers

Name	Address
Weaver & Co	565 Nucla Way Unit B Aurora, CO, 80011 United States Of America

Products

1 . Adhesive, <specify>

Product Type	Single Device Product	Effective Date	13/10/2020
GMDN	10034 Adhesive, <specify>		
Intended Purpose	Adhesive conductive paste used to hold electrodes in place on the patient's skin.		

Specific Conditions

No Specific Conditions included on Record

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/website-copyright.htm>.

Public Summary